

5

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 5

5

Lot Number:  14125043		Exp. Date: 2015-11	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
Gln	mmol/L	 4.36	 5.86
Glu	mmol/L	 1.79	 2.39
Gluc	g/L	 8.38	 12.38
Lac	g/L	 4.46	 5.96



200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800