










Assay Data Sheet - Expected Ranges
 BioProfile® FLEX Controls - LEVEL 4

Lot Number:  533508		Exp. Date: 2017-06	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
Gln	mmol/L	 0.72	 1.22
Glu	mmol/L	 0.24	 0.74
Gluc	g/L	 0.74	 1.24
Lac	g/L	 0.75	 1.25



200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800