










5

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 5

5

Lot Number:  511117		Exp. Date: 2016-10	
Parameter	Units	Expected Range	
		Lower Limit	Upper Limit
Gin	mmol/L	 4.41	 5.91
Glu	mmol/L	 1.79	 2.39
Gluc	g/L	 8.29	 12.29
Lac	g/L	 4.61	 6.11

nova
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800